

Display Report

Analysis Info

Analysis Name D:\Data\test\1103\bamating_17_01_212.d
Method 212.m
Sample Name bamating
Comment

Operator
Instrument bruker
amaZon SL

Acquisition Parameter

Ion Source Type	ESI	Ion Polarity	Positive	Alternating Ion Polarity	on
Mass Range Mode	UltraScan	Scan Begin	70 m/z	Scan End	1000 m/z
Capillary Exit	140.0 V	n/a	n/a	Trap Drive	50.6
Accumulation Time	1405 μ s	Averages	5 Spectra	Auto MS/MS	off

